
OxisResearch™

A Division of OXIS Health Products, Inc.

BIOXYTECH® Urinary 8-epi-Prostaglandin F_{2α}

Enzyme Immunoassay for Urinary Isoprostane

For Research Use Only. Not For Use In Diagnostic Procedures.

Catalog Number 21048 This product is patent protected.

INTRODUCTION

The Analyte

Isoprostanes are prostaglandin-like compounds that are produced by peroxidation of lipoproteins (1,2). 8-epi-prostaglandin-F_{2α} has been shown to be a potent vasoconstrictor in rat kidneys (3) and rabbit lungs (4). Isoprostanes may also play a role in atherosclerosis (5,6). Measurement of isoprostanes concentration may be helpful in assessment of oxidative stress, hepatorenal syndrome, rheumatoid arthritis, atherosclerosis and carcinogenesis (7). This kit can be used for the quantitation of free 8-epi-prostaglandin-F_{2α} in urine samples without the need for prior purification or extraction.

Principles of the Procedure

The BIOXYTECH® Urinary 8-epi-Prostaglandin F_{2α} Assay is a competitive enzyme-linked immunoassay (ELISA) for determining levels of 8-epi-prostaglandin-F_{2α} in urine samples. Briefly, the samples are mixed with an enhancing reagent that essentially eliminates interferences due to non-specific binding. The 8-epi-prostaglandin-F_{2α} in the sample or standard then competes with 8-epi-prostaglandin-F_{2α} conjugated to horseradish peroxidase (HRP Conjugate) for binding to a polyclonal antibody specific for 8-epi-prostaglandin-F_{2α} coated on the microplate. Following substrate addition, the intensity of the color is inversely proportional to the amount of unconjugated 8-epi-prostaglandin-F_{2α} in the sample or standard.

REAGENTS

Materials Provided

- | | |
|--|------------|
| • 96-well microtiter plate, pre-coated with 8-epi-prostaglandin-F _{2α} antibody | 1 |
| • 8-epi-prostaglandin-F _{2α} Standards (1µg/mL) | 2 x 60 µL |
| • Enhanced Dilution Buffer | 100 mL |
| • Wash Buffer | (5x) 40 mL |
| • TMB Substrate (Tetramethylbenzidine) | 25 mL |
| • HRP Conjugate | 250 µL |
| • Disposable reagent troughs for a multichannel pipettor | 2 |
| • ELISA Template | 1 |
| • Stop Solution | 5 mL |

Materials Required But Not Provided

- Precision pipettes with disposable tips. A multichannel pipette is helpful, but not required.
- 96-well microplate reader for measurement of absorbance at 450 nm.
- Reagents for the quantification of creatinine for normalization.
- Deionized water.

Warnings and Precautions

- Do not smoke, eat or drink in areas where samples and reagents are handled.
- Wear disposable gloves when handling samples and reagents.
- Do not pipette reagents or samples by mouth.
- In case of accidental exposure of skin, mucous membranes or eyes to the components of this kit, thoroughly wash the exposed area with water.
- For *in vitro* use only. For research purposes only. Not for use in diagnostic procedures.

Reagent Storage and Handling

Store all components at 4°C until immediately before use. Do not freeze.

PROCEDURE

Note: This product is intended for use with urine samples and has not been validated for use with serum, tissue culture supernatants or tissue extracts.

The following instructions are based on using the entire kit (all of the wells at one time). If portions of the kit are to be used at a later time, one may desire to prepare smaller quantities and save the remaining stock for later use.

Reagent Preparation

1. Substrate and Enhanced Dilution Buffer: Ready to use.
2. Add the 5x Wash Buffer (40 mL) to 160 ml of deionized water, mix well.
3. HRP Conjugate:
 - a. Centrifuge vial before removing the cap.
 - b. Prepare a 1/50 dilution of the HRP conjugate of sufficient volume for the amount of samples to be run that day. For the performance of the entire assay, add 240 µL of conjugate to 11.760 ml of Enhanced Dilution Buffer.

Preparation of Standards

Prepare a series of standards by diluting the 8-epi-prostaglandin-F_{2α} Standards (1µg/ml) to the following concentrations: 100, 50, 10, 5, 1, 0.1 and 0.05 ng/ml

- S7: Add 50 µL of Standard to 450µL of Dilution buffer (1x) = 100 ng/ml
- S6: Add 200 µL of S7 to 200 µL of Dilution Buffer (1x) = 50 ng/ml
- S5: Add 100 µL of S6 to 400 µL of Dilution Buffer (1x) = 10 ng/ml
- S4: Add 200 µL of S5 to 200 µL of Dilution Buffer (1x) = 5 ng/ml
- S3: Add 100 µL of S4 to 400 µL of Dilution Buffer (1x) = 1 ng/ml
- S2: Add 100 µL of S3 to 900 µL of Dilution Buffer (1x) = 0.1 ng/ml
- S1: Add 500 µL of S2 to 500 µL of Dilution Buffer (1x) = 0.05 ng/ml
- S0: Dilution Buffer (1x) only.

Sample Preparation

1. Make multiple dilutions of each sample, e.g. 1:2; 1:4; 1:8 using Dilution Buffer (1x). Optimal dilution will need to be determined by the researcher.


Assay

1. Remove microplate from foil pouch.
2. Pipet 100 μL of the prepared standard or sample mixture into each well.
3. Pipet 100 μL of diluted HRP Conjugate to each well, except for the reagent blank (S0). Add 100 μL of Dilution Buffer to blank, instead.
4. Seal plate or place in humidity chamber.
5. Allow the plate to stand at room temperature for 2 hours.
6. Empty the contents and blot the plate on a lint free towel.
7. Add 300 μL of diluted Wash Buffer (x1) and let stand 2-3 minutes.
8. Repeat steps 6 and 7 two additional times.
9. Empty the contents and blot the plate on a lint free towel.
10. Add 200 μL TMB Substrate to each well. Incubate for 20-40 minutes at room temperature, until a blue hue can be seen for the blank, S0.
11. Add 50 μL of Stop Solution to each well. The color in each well should change from blue to yellow. Read absorbance at 450 nm.

Calculations

An example of what the Standard Curve should look like is pictured below. The Standard Curve is obtained by fitting the Standard absorbances at 450 nm to the concentration of 8-epi-prostaglandin- $\text{F}_{2\alpha}$ by a logistic curve fit method. Average absorbance values obtained for the Blanks and subtract this value from all other wells.

Figure 1: 8-epi-Prostaglandin- $\text{F}_{2\alpha}$ Standard Curve



PERFORMANCE CHARACTERISTICS

Specificity*

8-epi-Prostaglandin F _{2α}	100.0%
9 _α ,11β-Prostaglandin F _{2α}	4.1%
13,14-Dihydro-15-Keto-PGF _{2α}	3.0%
9 _α ,11β-Prostaglandin F _{2α}	<0.01%
Prostaglandin F _{2α}	<0.01%
6-Keto-Prostaglandin F _{1α}	<0.01%
Prostaglandin E ₂	<0.01%
Prostaglandin D ₂	<0.01%
Arachidonic Acid	<0.01%

*Cross reactivity at mid-point of the Standard Curve.

REFERENCES

1. Roberts LJ, Morrow JD (1997). The generation and action of isoprostanes. *Biochim.Biophys.Acta.* **1345**: 121-135.
2. Morrow JD, Roberts LJ (1997). The isoprostanes: unique bioactive products of lipid peroxidation. *Prog. Lipid Res.* **36(1)**: 1-21.
3. Kanji VK, Wang C, Salahudeen AK (1999). Vitamin E suppresses cyclosporin A-induced increase in the urinary excretion of arachidonic acid metabolites inducing F2-isoprostanes in the rat model. *Transplant Proc.* **31(3)**: 1724-28.
4. Banerjee M, Kang KH, Morrow JD, Roberts LJ, Newman JH (1992). Effects of a novel prostaglandin 8-epi PGF₂ alpha in rabbit lung in situ. *Am. J. Physiol.* **263**: H660-H663.
5. Gniwotta C, Morrow JD, Roberts LJ II, Kuhn H (1997). Prostaglandin F₂-like compounds, F₂-isoprostanes, are present in increased amounts in human atherosclerotic lesion. *Arterioscler. Thromb. Vasc. Biol.* **17**: 3236-3241.
6. Tangirala RK et al. (2001). Reduction of isoprostanes and regression of advanced atherosclerosis by apolipoprotein E. *J. Biol. Chem.* **276(1)**: 261-6.
7. Roberts LJ and Morrow JD (2000). Measurement of F₂-isoprostanes as an index of oxidative stress *in vivo*. *Free Rad. Biol. Medicine.* **28(4)**: 505-513.

LIMITED WARRANTY

OXIS Health Products, Inc. warrants that, at the time of shipment, this product is free from defects in materials and workmanship. OXIS Health Products, Inc. makes no warranty, expressed or implied, including, but not limited to, the warranties of fitness for a particular use and merchantability which extends beyond the description of the product on the face hereof, except that this product will meet our specifications at the time of shipment.

OXIS Health Products, Inc. must be notified of any breach of this warranty within 14 days of receipt of this product. No claim shall be honored if OXIS Health Products, Inc. is not notified within this time period, or if the product has been stored in any way other than described in this product insert. The sole and exclusive remedy of the buyer for any liability based upon this warranty is limited to the replacement of this product, or refund of the invoice price of the goods. OXIS Health Products, Inc. shall not be liable otherwise or for incidental or consequential damages, including but not limited to the costs of handling.

TECHNICAL SUPPORT

An OXIS Health Products, Inc. Technical Support Representative can be reached by telephone at (800) 547-3686 or (650) 212-2568 Ext. 502, or by e-mail techsupport@oxis.com Monday through Friday 8:00 AM to 5:00 PM Pacific Time.

OxisResearch®
323 Vintage Park Drive, Suite B
Foster City, CA 94404-1136 U.S.A.
Telephone: 650-212-2568 or 800-547-3686
Fax: 650-212-2569
www.oxisresearch.com
Last Revision May 2006

Made in the U.S.A.
BIOXYTECH® is a registered trademark of OXIS International, Inc.
Foster City, CA 94404-1136
Copyright© 2002 OXIS Health Products, Inc. All rights reserved